

Radon Measurement Procedure

Prior to the Test

It is important to have closed house conditions begin a minimum of 12 hours prior to the radon measurement. A closed house is when all of the openings to the exterior are shut off except for normal entry and exit. The client meets with our licensed radon tester to go over the contract and give the client a copy of the contract. Payment for the radon measurement is due at this time. Our licensed radon tester is also available at this time to answer questions in regard to radon.

Radon Test

The Radalink monitor is placed in the lowest livable level of the home after the closed house condition has been established. Our licensed radon tester sets the monitor into testing mode via a Bluetooth computer program. The test is run continuously for the next 48 hours taking samples of the radon level, temperature and barometric pressure. Tilt sensors continuously monitor the device which will tell if the monitor has been moved to another location during the test.

Completion of the Test

After the completion of the test the data is retrieved from the monitor then uploaded to Radalink Laboratory for analysis. Within 24 hours the test results are sent to the client via a detailed report. Our inspector and Radalink are available to answer your questions in regard to the results of the test.